

Certificate of Compliance

| Certificate: | 80188655 | Master Contract: | 187869 |
|--------------|----------------------------------------------------------------------------------------------------------------------------------------|------------------|------------|
| Project: | 80219638 | Date Issued: | 2024-09-16 |
| Issued To: | Crystal Engineering Corp. 708 Fiero Lane, Suite 9 San Luis Obispo, California, 93401 United States Attention: Trevor Leger | | |

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: Jerry Thai Jerry Thai

PRODUCTS

CLASS - C225804 - PROCESS CONTROL EQUIPMENT - Intrinsically Safe Entity - For Hazardous Locations CLASS - C225884 - PROCESS CONTROL EQUIPMENT - Intrinsically Safe Entity - For Hazardous

Locations - Certified to US Standards

Class I, Division 1, Groups A, B, C and D T* Ex ia IIC T* Ga Class I, Zone 0, AEx ia IIC T* Ga

Digital Test Gauge Model Series XP3i and XP3i-DD; Intrinsically safe when powered by 3 primary 1.5V AA cells(Approved cell type only*); USB port: Um = 6 V; Ambient temperature various**. Maximum working pressure: up to 15,000 PSI.

*Note: refer to Conditions of Acceptability for temperature class and ambient temperature.



Certificate: 80188655 **Project:** 80188655 Master Contract: 187869 Date Issued: 2024-09-16

Conditions of Acceptability:

- 1. Enclosure of the XP3i and XP3i-DD is made of aluminum, if it is mounted in an area where the use of EPL Ga apparatus is required, it must be installed such that, even in the event of rare incidents, ignition sources due to impact and friction sparks are excluded.
- 2. To avoid sparks caused by electrostatic discharge, the XP3i must be earthed during use in hazardous locations. This may be accomplished by providing an appropriate and continuous path to earth through the pressure fitting, the metal enclosure or the hand of the user.
- 3. USB connector shall not be used in Hazardous Locations.
- 4. Ambient temperature range and Temperature class depends on the type of the used cells:

| Approved Cell Type* | Ambient temperature** | Temperature class |
|---------------------------|-----------------------|-------------------|
| Rayovac 815 | -20 to 50 °C | |
| Duracell MN1500 | -20 to 50 °C | |
| Energizer E91 | -18 to 50 °C | T2C |
| Energizer EN91 | -18 to 50 °C | 150 |
| Panasonic LR6XWA | -20 to 45 °C | |
| Varta 4106 Longlife | -10 to 45 °C | |
| Varta 4706 Max Tech | -20 to 45 °C | |
| Energizer EN91 Industrial | -18 to 45 °C | T4 |
| Varta 4906 Longlife Power | -10 to 45 °C | |

APPLICABLE REQUIREMENTS

| CAN/CSA C22.2 No. 61010-1-12 + | Safety Requirements for Electrical Equipment for Measurement, |
|--------------------------------|---------------------------------------------------------------|
| UPD1:2015, UPD2:2016, AMD 1-18 | Control, and Laboratory Use — |
| (<i>R2022</i>) | Part 1: General Requirements |
| CSA C22.2 No. 60079-0:19 | Explosive atmospheres – |
| | Part 0: Equipment – General requirements |
| CAN/CSA C22.2 No. 60079-11:14 | Explosive atmospheres – |
| (R2018) | Part 11: Equipment protection by intrinsic safety "i" |
| ANSI/UL 61010-1 | Safety Requirements for Electrical Equipment for Measurement, |
| 3rd Edition (2012), AMD1 :2018 | Control, and Laboratory Use — |
| | Part 1: General Requirements |
| ANSI/UL 60079-0-2020 | Explosive atmospheres – |
| Seventh Edition | Part 0: Equipment – General requirements |
| ANSI/UL 60079-11-2018 | Explosive Atmospheres – |
| Sixth Edition | Part 11: Equipment Protection by Intrinsic Safety "i" |



Certificate: 80188655 **Project:** 80188655 Master Contract: 187869 Date Issued: 2024-09-16

MARKINGS

The manufacturer is required to apply the following markings:

- Products shall be marked with the markings specified by the particular product standard.
- Products certified for Canada shall have all Caution and Warning markings in both English and French.

Additional bilingual markings not covered by the product standard(s) may be required by the Authorities Having Jurisdiction. It is the responsibility of the manufacturer to provide and apply these additional markings, where applicable, in accordance with the requirements of those authorities.

The products listed are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US (indicating that products have been manufactured to the requirements of both Canadian and U.S. Standards) or with adjacent indicator 'US' for US only or without either indicator for Canada only.

The following markings are provided:

- Manufacturer's name: "Crystal Engineering Corp.", or CSA Master Contract Number "187869", adjacent to the CSA Mark in lieu of manufacturer's name.
- Model designation: As specified in the PRODUCTS section, above.
- Electrical ratings: As specified in the PRODUCTS section, above.
- Ambient temperature rating: As specified in the PRODUCTS section, above.
- Manufacturing date, or serial number, traceable to year and month of manufacture.
- The CSA Mark, as shown on the Certificate of Compliance.
- The designation "CSA24CA80188655X".
- Hazardous Location designation: As specified in the PRODUCTS section, above. The word "Class" may be abbreviated "CL", the word "Division" may be abbreviated "DIV", and the word "Groups" may be abbreviated "GRP" or "GP".
- Method of Protection markings (Ex markings): As specified in the PRODUCTS section, above. The word "Class" may be abbreviated "CL", the word "Zone" may be abbreviated "ZN".
- Temperature code: As specified in the PRODUCTS section, above.
- The following words, or suitable equivalent, in English and French:
 - "Refer to Operation Manual" and "Reportez-vous au Mode d'emploi".
- The following words, or suitable equivalent, in English and French, on the instruction manual:
 - For Canada: "INTRINSICALLY SAFE" and "SECURITE INTRINSEQUE", or "IS" or "I.S." or the symbol "Ex ia"
 - "WARNING: SUBSTITUTION OF COMPONENTS MAY IMPAIR INTRINSIC SAFETY", and "AVERTISSEMENT: LA SUBSTITUTION DE COMPOSANTS PEUT COMPROMETTRE LA SECURITE INTRINSEQUE"
 - Where $Um \neq 250V$: The words "Maximum non-hazardous voltage not to exceed 6V."
 - "REPLACE 3 AA BATTERIES WITH APPROVED TYPES IN NON-HAZARDOUS LOCATIONS ONLY" and "REMPLACEZ 3 PILES AA PAR DES TYPES APPROUVÉS UNIQUEMENT DANS DES ENDROITS NON DANGEREUX"



Certificate: 80188655 **Project:** 80188655 Master Contract: 187869 Date Issued: 2024-09-16

Notes:

Products certified under Class C225804 C225884 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80188655

Master Contract: 187869

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

| Project | Date | Description |
|----------|------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 80219638 | 2024-09-16 | Evaluation to update Report 80188655 for new P/N 6356-7 for use up to 15,000 PSI. |
| 80188655 | 2024-01-30 | Original cCSAus certification of intrinsically safe XP3i Series Digital Pressure Gauge based on the acceptance of IECEx Certificate IECEx FTZU 23.0017X. |
| | | Markings: IS Class I, Division 1, Groups A, B, C, D T3CT4 Ex ia IIC Ga Class I, Zone 0 AEx ia IIC Ga -20°C to +50°C |